

Test Results

S9110 Windscreen Retention Test 6 Hour Soak

Customer: Thomas KÜGEL
 PETEC Verbindungstechnik GmbH,
 Wüstenbuch 26, 96132
 SCHLÜSSELFELD,
 GERMANY

Contact: Paul Moore
 HORIBA MIRA Ltd
 Safety Development Dept
 Watling Street
 Nuneaton
 Warwickshire
 CV10 0TU, UK
 +44(0)24 7635 5000

Authority: S84925

Witnesses: Rudolf Gerlach - TUV Rheinland

Test Date(s) 30 November 2017

Test Objective / Method / Specification No

To assess PETEC windscreen bonding product performance according to FMVSS212. Vehicle was soaked at 23°C and 50% relative humidity. The test was conducted 6 hours after the windscreen was fitted. MIRA Test Number S9110.

Specimen Description / Part No(s)

PETEC Scheibenkleber (83600), PETEC Multiaktiv Primer (82410), PETEC Aktivator (82230).

Test vehicle:

Make Ford
 Model Mondeo
 Drive hand RH drive
 VIN WF05XXGBB57C13161

Test Results Summary

Results only relate to items tested. The subject was tested in accordance with the test specification with without deviation. The acceptance criteria of the test specification were:	Met	✓	Not Met	See comments
--	-----	---	---------	--------------

Prepared By:



William Martin
 Senior Engineer - Crash

Approved By:



Colin Smith
 Head of Crash



Date: 29/09/21

© HORIBA MIRA Ltd 2021. All rights reserved, subject to client contract. Information contained in this document may not be published in any form of advertising or other matter without prior agreement of the CEO of HORIBA MIRA.

HORIBA MIRA Ltd. Registered Office: Watling Street · Nuneaton · Warwickshire · CV10 0TU · England · <http://www.horiba-mira.com>
 Tel: +44 (0)24 7635 5000 · Fax: +44 (0)24 7635 8000 Registered in England No 9626352 · VAT Registration GB 100 1464 84

Tests marked "Not UKAS Accredited" in this report are not included in the UKAS Accreditation Schedule for this laboratory. Opinions and interpretations expressed herein are outside the scope of UKAS Accreditation

Product Details

Part Description	Part Number	Date Received
Primer	PETEC 82410	27 & 28/11/17
Activator	PETEC 82230	27 & 28/11/17
Sealant	PETEC 83600	27 & 28/11/17

Test Results Detail

The test vehicle was soaked at a 6 hour average temperature of 23°C and relative humidity 50% after windscreen installation, which was carried out by 24 Screen Savers Ltd. It was then subjected to a 100% frontal impact by being propelled into a rigid crash block at 30mph (48km/h), as described in FMVSS212 Section 5. The ATDs were restrained by the vehicle's standard seat belts and the airbags triggered by the vehicle standard system.

Test Conditions				
Tyre Pressure (bar)	Front	2.1	Rear	2.1
Vehicle modifications	None			
Steering Column	Mid position			
Windows	Down			
Seatbelt Height Adjust	Mid position			
Doors	Unlocked			
Parking Brake	Off			
Ignition	On			
Seat Adjustment - Driver	Mid Fore/Aft – Lowest Height			
Seat adjustment – Passenger	Mid Fore/Aft – Lowest Height			
Restraint system specification	Driver and passenger airbag			

Vehicle mass details	Front (kg)	Rear (kg)	Total (kg)
Unloaded vehicle mass	792	555.5	1347.5
Test Weight (in test condition, including 93% fuel, nominal fluids, instrumentation, ballast and 2 occupants)	865	829.5	1694.5

Assessment against Legislative Criteria		
Impact Velocity (Target 48.3 +1 / -0 km/h)	48.6 km/h	Complied
Impact Alignment (target <5°. Approx. ± 235mm for 2700mm wheelbase)	45 mm left	Complied
Performance assessment: (Max 25% detachment on each side of windscreen perimeter)	LH perimeter 0%	Complied
	RH perimeter 0%	Complied

Test Equipment

Rigid Barrier with plywood facing

2x Hybrid III 50%ile ATDs (Anthropomorphic Test Devices) – un-instrumented for ballast only

Measurement equipment as listed in Appendix 6

6 high speed digital cameras

Attachments

Appendix 1 - Test Photographs

Appendix 2 - Quality Assurance of Measurements

Appendix 1 Test Photographs



Photo 1
LH General view – Pre-Test



Photo 2
LH General view – Post-Test



Photo 3
Front view – Pre-Test



Photo 4
Front view – Post-Test



Photo 5
RH General view – Pre-Test



Photo 6
RH General view – Post-Test



Photo 7
Close front view of LH A-pillar / windscreen LH edge – Post-Test



Photo 8
Close front view of RH A-pillar / windscreen RH edge – Post-Test



Photo 9
Close front view of header rail / windscreen top edge – Post-Test



Photo 10
Close front view of scuttle / windscreen lower edge – Post-Test



Photo 11
Post-test LHF ATD side view – Post-Test



Photo 12
Post-test RHF ATD side view – Post-Test

Appendix 2 Quality Assurance of Measurements

All instrumentation, high speed images and associated analysis contained in this report conforms to the requirements within SAE J211 July 2007.

The test equipment is checked on a regular schedule to traceable standards in an International Assurance of Measurements (QAM) procedure. Each item of equipment is issued with a QAM number.

The numbers for the equipment used in these tests were:-

Item	QAM number	Cal due date
ATD Identification – LHF HIII 50% No 75	N/A	N/A
ATD Identification – RHF HIII 50% No 138	N/A	N/A
Weigh Scales	38720-23 38724-27	29/06/2018 20/08/2018
5m Steel Tape Measure	34848	07/11/2022
Digital Level	33520	18/01/2018
Stop Watch	34851 38460	15/12/2018 17/05/2018
Tyre Pressure gauge	39679	09/03/2018
Impact Speed Measure (fixed)	8167	26/02/2018
Impact speed measure (mobile)	17921	29/07/2018
Climatic Control Temperature Probe	39704-07, 39785-88	03/01/2018
Climatic Control Humidity Probe	31995-96 31997-98	18/08/2018 19/01/2018
32ch Thermocouple Amplifier	30279	03/01/2018
Climactic Chamber Controller	34000	05/01/2018

Camera ID	View	QAM number	Cal due date
1	F01 LH View – Whole Vehicle	37921	01/08/2018
2	F02 LH View – Front Half of Vehicle	37924	13/07/2018
3	F06 RH View – Whole Vehicle	37938	17/08/2018
4	F07 RH View – Front Half of Vehicle	37922	22/11/2018
5	F14 Overhead View – Whole Vehicle	37919	18/07/2018
6	F11 Front View – Front Half of Vehicle	37917	18/07/2018
7			
8			

Channel Sample Rate: N/A

Weigh Scales

CALIBRATION CERTIFICATE

Issued by **HORIBA MIRA Ltd**

Issue date: 29 Jun 2017 Cert No : 33720170617



Page 1 of 5 Pages

Approved Signatory

P. Macleod

Walling Street, Nuneaton
Warwickshire, CV10 0TU, UK.
Telephone : +44 (0)24 7635 5225
Facsimile : +44 (0)24 7635 8225
<http://www.horiba-mira.com>

- Mark Pickering – Department Manager
- Philip Macleod – Supervisor – Instrument Calibration
- Dominic Mhandu – Metrologist

Client:	Test Operations FG	Manufacturer:	Intercomp
Section:	Safety, Crash Off-Board instruments	Model:	170127-WPC
Address:	Walling Street Nuneaton Warwickshire	Description:	Weigh Pads
	--	Serial No:	0216MC15008
Client ID:	--	Calibration Date:	29 Jun 2017
MIRA ID:	Q38720 to Q38723	Calibration Procedure:	QA4299/C/07
Date received:	17 Jun 2017	Equipment used:	Page 2-5
Dallas ID:		Measurement Results:	Page 2-5
		Measurement Uncertainty:	Page 2-5

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5 Hz with a total harmonic distortion of less than 3%.

This instrument was calibrated by comparison with force measurement reference standards using a MIRA procedure. Where the instrument has an electrical output and this has been calibrated, the output has been measured using electrical reference standards.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd. Registered office: Walling Street, Nuneaton, Warwickshire, CV10 0TU. Registered in England No: 5526352. VAT Registration number: GB100145484

TC0032b Issue 20

CALIBRATION CERTIFICATE
Issued by HORIBA MIRA Ltd
Issue date: 20 Aug 2017 Cert No : 38724170817



Page 1 of 5 Pages

Approved Signatory



Walling Street, Nuneaton
 Warwickshire, CV10 0TU, UK.
 Telephone : +44 (0)24 7635 5225
 Facsimile : +44 (0)24 7635 8225
<http://www.horiba-mira.com>

□ Mark Pickering – Department Manager
 □ Philip Macleod – Supervisor – Instrument Calibration
 □ Dominic Mhardu – Metrologist

Client:	Test Operations FG	Manufacturer:	Intercomp
Section:	Safety, Crash Off-Board	Model:	170127-WPC
Address:	Walling Street Nuneaton Warwickshire	Description:	Weight Pads
Client ID:	--	Serial No:	0216MC15003
MIRA ID:	Q38724 to Q38727	Calibration Date:	20 Aug 2017
Date received:	17 Aug 2017	Calibration Procedure:	QA4299/CIC7
Dallas ID:		Equipment used:	Pages 2 to 5
		Measurement Results:	Pages 2 to 5
		Measurement Uncertainty:	Pages 2 to 5

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5Hz with a total harmonic distortion of less than 3%.

This instrument was calibrated by comparison with force measurement reference standards using a MIRA procedure. Where the instrument has an electrical output and this has been calibrated, the output has been measured using electrical reference standards.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Walling Street, Nuneaton, Warwickshire, CV10 0TU. Registered in England No: 3525352. VAT Registration number: GB 00146484.

TC002b Issue 20

5m Steel Tape Measure

CALIBRATION CERTIFICATE

Issued by **HORIBA MIRA Ltd**

Issue date: 08 Nov 2017 Cert No : 34648181017



Page 1 of 2 Pages

Approved Signatory

P. Macleod

Watling Street, Nuneaton
Warwickshire, CV11 0JU, UK.
Telephone : +44 (0)24 7635 5225
Facsimile : +44 (0)24 7635 8226
<http://www.horiba-mira.com>

- Mark Pickering – Department Manager
- Philip Macleod – Supervisor – Instrument Calibration
- Dominic Mhandu – Metrologist

Client:	Test Operations FG	Manufacturer:	Assist
Section:	Safety, Crash Off-Board Instruments	Model:	32G-5019
Address:	Watling Street	Description:	5m Steel Tape Measure
	Nuneaton	Serial No:	--
	Warwickshire	Calibration Date:	08 Nov 2017
	--	Calibration Procedure:	QA31C5/C/03
Client ID:	--	Equipment used:	Page 2
MIRA ID:	Q34648	Measurement Results:	Page 2
Date received:	18 Oct 2017	Measurement Uncertainty:	Page 2
Dallas ID:			

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.6Hz with a total harmonic distortion of less than 3%.

This instrument was calibrated by comparison with length measurement reference standards using a MIRA procedure which incorporates limits based on the tolerances contained in document NIST handbook 44 section 5.52.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Watling Street, Nuneaton, Warwickshire, CV11 0JU. Registered in England No: 3628362. VAT Registration number: GB100146484

TC0032b Iss:0 20

Digital Level

CALIBRATION CERTIFICATE

Issued by HORIBA MIRA Ltd

Issue date: 18 Jan 2017 **Cert No :** 33520130117



Page 1 of 2 Pages

Approved Signatory



Walling Street, Nuneaton
 Warwickshire, CV10 0TU, UK.
 Telephone : +44 (0)24 7635 5225
 Facsimile : +44 (0)24 7635 8225
<http://www.horiba-mira.com>

|| Mark Pickering – Head of Instrument Calibration & Repair
 || Philip Macleod – Supervisor – Instrument Calibration
 || Dominis Mhandu – Metrologist

Client:	Test Operations FG	Manufacturer:	Smartool
Section:	Safety, Crash Off-Board Instruments	Model:	--
Address:	Walling Street Nuneaton Warwickshire --	Description:	Digital Level
Client ID:	--	Serial No:	--
MIRA ID:	Q33520	Calibration Date:	18 Jan 2017
Date received:	13 Jan 2017	Calibration Procedure:	QA3129/C/03
Dallas ID:		Equipment used:	Page 2
		Measurement Results:	Page 2
		Measurement Uncertainty:	Page 2

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5Hz with a total harmonic distortion of less than 3%.

Instrument calibrated by comparison with angular reference standards using MIRA Procedure QA3129/C.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Walling Street, Nuneaton, Warwickshire, CV10 0TU. Register in England No: 3628352. VAT Registration number: GB100146464.

1.00002b Issue 20

Stop Watch

CALIBRATION CERTIFICATE

Issued by MIRA Ltd

Issue date: 18 May 2015 **Cert No :** 38460230315



Page 1 of 2 Pages

Approved Signatory

P. Macleod

Walling Street, Nuneaton
 Warwickshire, CV10 0TU, UK
 Telephone : +44 (0)24 7635 5225
 Facsimile : +44 (0)24 7635 8225
<http://calibration.mira.co.uk>

Mark Pickering – Head of Instrument Calibration & Repair
 Philip Macleod – Supervisor – Instrument Calibration

Client:	MIRA, Test Operations FG	Manufacturer:	RS
Section:	Vehicle Env & Aero, Climatic Chamber	Model:	811-1818
Address:	Walling Street Nuneaton Warwickshire	Description:	Stop Watch
Client ID:	--	Serial No:	--
MIRA ID:	Q38460	Calibration Date:	18 May 2015
Date received:	23 Mar 2015	Calibration Procedure:	QA2113/C/02
		Equipment used:	Page 2
		Measurement Results:	Page 2
		Measurement Uncertainty:	Page 2

Condition of Instrument: New

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5Hz with a total harmonic distortion of less than 3%.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2015 MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the Managing Director of MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Walling Street, Nuneaton, Warwickshire, CV10 0TU. Registered in England No: 402570. VAT Registration number: GB100148484.

TC0032b Issue 17

CALIBRATION CERTIFICATE
Issued by HORIBA MIRA Ltd
Issue date: 16 Dec 2015 **Cert No :** 34851251115



Page 1 of 2 Pages

Approved Signatory

P. MacLeod

Walling Street, Nuneaton
 Warwickshire, CV19 0TU, UK.
 Telephone : +44 (0)24 7635 5225
 Facsimile : +44 (0)24 7635 8225
<http://www.horiba-mira.com>

- || Mark Pickering – Head of Instrument Calibration & Repair
- || Philip Macleod – Supervisor – Instrument Calibration
- || Miroslaw Palucki – Metrologist

Client:	Mechanical Engineering FG	Manufacturer:	RS
Section:	Braking	Model:	639-9259
Address:	Walling Street Nuneaton Warwickshire	Description:	Digital Stopwatch
	--	Serial No:	--
Client ID:	--	Calibration Date:	15 Dec 2015
MIRA ID:	Q54851	Calibration Procedure:	CA2113/C/02
Date received:	25 Nov 2015	Equipment used:	Page 2
Dallas ID:		Measurement Results:	Page 2
		Measurement Uncertainty:	Page 2

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5Hz with a total harmonic distortion of less than 3%.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2015 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Walling Street, Nuneaton, Warwickshire, CV19 0TU. Registered in England No: 9625352. VAT Registration number: GB102146484.

TC0032b Issue 17

Tyre Pressure Gauge

CALIBRATION CERTIFICATE


Issued by HORIBA MIRA Ltd

Issue date: 09 Mar 2017 Cert No : 38679-00117



Page 1 of 3 Pages

Approved Signatory

P Macleod 

Walling Street, Nuneaton
Warwickshire, CV10 0TU, UK.
Telephone : +44 (0)24 7635 5225
Facsimile : +44 (0)24 7635 8225
<http://www.horiba-mira.com>

- Mark Pickering - Head of Instrument Calibration & Repair
- Philip Macleod - Supervisor - Instrument Calibration
- Dominic Mhandu - Metrologist

Client:	Test Operations FG	Manufacturer:	PCL
Section:	Safety, Crash Off-Board Instruments	Model:	AFG1H03
Address:	Walling Street Nuneaton Warwickshire --	Description:	12 bar Tyre Inflator
Client ID:	--	Serial No:	141121095
MIRA ID:	Q33679	Calibration Date:	09 Mar 2017
Date received:	10 Jan 2017	Calibration Procedure:	QA4097/C/03
Dallas ID:		Equipment used:	Page 3
		Measurement Results:	Page 2
		Measurement Uncertainty:	Page 2

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.6Hz with a total harmonic distortion of less than 3%.

This Instrument was calibrated by comparison with pressure measurement reference standards using a MIRA procedure which incorporates limits based on the applicable standard, BS EN 12645:1999.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Walling Street, Nuneaton, Warwickshire, CV10 0TU. Registered in England No: 9826352. VAT Registration number: GB100146484

TC0052: Issue 20

9 Beam Speed Measurement Laser

CALIBRATION CERTIFICATE

Issued by HORIBA MIRA Ltd

Issue date: 19 Apr 2017 **Cert No :** 08167190417



Page 1 of 5 Pages

Approved Signatory



Watling Street, Nuneaton
Warwickshire, CV10 0TU, UK.
Telephone : +44 (0)24 7635 6225
Facsimile : +44 (0)24 7635 8225
http://www.horiba-mira.com

- Mark Pickering – Head of Instrument Calibration & Repair
- Philip Macleod – Supervisor – Instrument Calibration
- Dominic Mhandu - Metrologist

Client:	Test Operations FC	Manufacturer:	MIRA
Section:	Safety, Crash Off-Board Instruments	Model:	--
Address:	Watling Street	Description:	9 Beam Speed Measurement Sys
	Nuneaton	Serial No:	--
	Warwickshire	Calibration Date:	20 Apr 2017
	--	Calibration Procedure:	QA2364/C/05
Client ID:	--	Equipment used:	Page 5
MIRA ID:	Q08167	Measurement Results:	Pages 2 to 4
Date received:	19 Apr 2017	Measurement Uncertainty:	Page 5
Dallas ID:			

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	No
Adjusted during calibration	No
Repaired prior to or during calibration	Yes
Within specification on completion, at the points measured subject to the measurement uncertainty	No

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5Hz with a total harmonic distortion of less than 3%.

This speed measurement system was calibrated by comparison with distance and time measurement reference standards using a MIRA procedure which incorporates limits based on client requirements detailed in request for service document CR225072012A.

This calibration certificate includes the laser speed measurements taken after repair.

Laser speed set C1 still does not conform to specification.

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.



Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Watling Street, Nuneaton, Warwickshire, CV10 0TU. Registered in England, No. 8526352. VAT Registration number: GB 130146484

TC0032b Issue 20

Mobile Speed Measurement Laser

CALIBRATION CERTIFICATE

Issued by HORIBA MIRA Ltd

Issue date: 21 Jun 2017 **Cert No :** 178992503171



Page 1 of 2 Pages

Approved Signatory

P. Macleod

Watling Street, Nuneaton
 Warwickshire, CV10 0JU, UK
 Telephone : +44 (0)24 7635 5225
 Facsimile : +44 (0)24 7635 8225
 http://www.horiba-mira.com

- Mark Pickering - Department Manager
- Philip Macleod - Supervisor - Instrument Calibration
- Dominic Mhanda - Metrologist

Client:	Test Operations FG	Manufacturer:	MIRA
Section:	Safety Crash Off-Board Instruments	Model:	TD590
Address:	Watling Street	Description:	Voie Speed Measurement Unit, Cable & Frame
	Nuneaton Warwickshire	Serial No:	--
Client ID:	--	Calibration Date:	04 Apr 2017
MIRA ID:	Q17899 to Q17901	Calibration Procedure:	QA2517/C/03
Date received:	25 Mar 2017	Equipment used:	Page 2
Dallas ID:		Measurement Results:	Page 2
		Measurement Uncertainty:	Page 2

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5Hz with a total harmonic distortion of less than 3%.

This instrument was calibrated by comparison with time and displacement measurement reference standards using a MIRA procedure which incorporates limits based on client requirements. These requirements are specified in document CR225072012B.

This certificate includes all component parts list.
 This certificate is a replacement for Certificate number 17899250317.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Watling Street, Nuneaton, Warwickshire, CV10 0JU. Registered in England No: 9626352. VAT Registration number: GB100146484.

TC0302b Issue 20

Climatic Temperature Probes

CERTIFICATE OF CALIBRATION

ISSUED BY: **Universal Instrument Services Ltd.**

Date of Calibration: 3rd January 2017
 Date of Issue: 5th January 2017

Certificate Number: 00048454



0125



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 69 The Whittle Estate, Cambridge Road,
 Whetstone, Leicester LE8 6PA
 Tel: 0116 275 0123 Fax: 0116 275 0262
 Website: www.uiscal.com
 Email: sales@uiscal.co.uk

Page 1 of 2 Pages
 Approved Signatory

J. Bruce
 J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUKEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 561434

Manufacturer: Not Known
Description: T⁺ Thermocouple
Model: T-TYPE
Serial No: Q39704
Asset No: Q39704
Order No: 263520
Date Received: 18th December 2016

CONDITION OF UNIT UNDER TEST

The Thermocouple was visually inspected prior to calibration

YES/NO
 Y

ADDITIONAL COMMENTS

STABILITY

The readings given are the results at the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate.

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page.

PROCEDURE

UIS procedure CP7.5.3

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. This uncertainty evaluation has been carried out in accordance with UKAS requirements.

TEST EQUIPMENT USED

Asset No.	Description	Certificate No.	Expiry date
ID3106	Hart 1590 Super Thermometer II	296385	21/MAR/2017
ID3051	Tinsley 81d Resistor 25 ohms	UKAS 0361304	09/11/2018
ID3032	PRT (25 ohms)	UKAS 47039	14/JUL/2017
ID3269	PRT (25 ohms)	UKAS 47042	14/JUL/2017
ID3243	PRT (25 ohms)	UKAS 47018	12/JUL/2017
D3245	PRT (25 ohms)	UKAS 47020	12/JUL/2017
D3276	Fluke 1586-2588 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: **Universal Instrument Services Ltd.**

Date of Calibration: 3rd January 2017
 Date of Issue: 5th January 2017
 Certificate Number: 00048455



0125

Universal



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 69 The Whittle Estate, Cambridge Road,
 Whetstone, Leicester LE8 5PA
 Tel: 0116 275 0123 Fax: 0116 275 0262
 Website: www.uiscal.com
 Email: sales@uiscal.co.uk

Page 1 of 2 Pages
 Approved Signatory

J. Bruce
 J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 331435

Manufacturer: Not Known
Description: 'T' Thermocouple
Model: T-TYPE
Serial No: Q39705
Asset No: Q39705
Order No: 265528
Date Received: 19th December 2016

CONDITION OF UNIT UNDER TEST

The Thermocouple was visually inspected prior to calibration

YES/NO

Y

ADDITIONAL COMMENTS

STABILITY

The readings given are the results at the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate.

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page.

PROCEDURE

UIS procedure CP7 5.3

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

TEST EQUIPMENT USED

<u>Asset No.</u>	<u>Description</u>	<u>Certificate No.</u>	<u>Expiry date</u>
IC3156	Hart 1690 Super-Thermometer II	258315	21/MAR/2017
IC3051	Tinsley Std Resistor 25 ohms	UKAS 0391304	09/JUL/2018
IC3032	PRT (25 ohms)	UKAS 47036	14/JUL/2017
IC3289	PRT (25 ohms)	UKAS 47042	14/JUL/2017
IC3240	PRT (25 ohms)	UKAS 47018	12/JUL/2017
IC3245	PRT (25 ohms)	UKAS 47020	12/JUL/2017
IC3276	Fluke 1568-2588 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATIONISSUED BY: **Universal Instrument Services Ltd.**Date of Calibration: 3rd January 2017
Date of Issue: 6th January 2017

Certificate Number: 00046450



0125



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 69 The Whittle Estate, Cambridge Road,
 Whetstone, Leicesters LE8 6PA
 Tel: 0116 275 0125 Fax: 0116 275 0382
 Website: www.uiscal.com
 Email: sales@uiscal.co.uk

Page 1 of 2 Pages
Approved Signatory

J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 331436

Manufacturer: Not Known
Description: T1 Thermocouple
Model: T-TYPE
Serial No: Q39700
Asset No: Q30706
Order No: 263528
Date Received: 18th December 2016

CONDITION OF UNIT UNDER TEST**YES/NO**

The Thermocouple was visually inspected prior to calibration.

Y

ADDITIONAL COMMENTS**STABILITY**

The readings given are the results of the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate.

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page.

PROCEDURE

UIS procedure CP7.5.3

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

TEST EQUIPMENT USED

Asset No.	Description	Certificate No.	Expiry date
U3156	Hart 1590 Super-Thermometer II	296385	21/MAR/2017
ID3051	Tinsley Std Resistor 25 ohms	UKAS 0391304	09/JUL/2016
ID3032	PRT (25 ohms)	UKAS 47039	14/JUL/2017
ID3260	PRT (25 ohms)	UKAS 47042	14/JUL/2017
ID3240	PRT (25 ohms)	UKAS 47018	12/JUL/2017
ID3245	PRT (25 ohms)	UKAS 47020	12/JUL/2017
IC3276	Fluke 1586-2588 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. The certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: **Universal Instrument Services Ltd.**

Date of Calibration: 3rd January 2017
 Date of Issue: 5th January 2017

Certificate Number: 00018457



0125



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 69 The Whittle Estate, Cambridge Road,
 Whetstone, Leicestershire LE8 6PA
 Tel: 0116 275 0123 Fax: 0116 275 0262
 Website: www.uscal.com
 Email: sales@uscal.co.uk

Page 1 of 2 Pages
 Approved Signatory

J. Bruce
 J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 351437

Manufacturer: Not Known
Description: 'T' Thermocouple
Model: T-TYPE
Serial No.: Q39707
Asset No.: Q39707
Order No.: 263520
Date Received: 19th December 2016

CONDITION OF UNIT UNDER TEST

The Thermocouple was visually inspected prior to calibration YES/NO
Y

ADDITIONAL COMMENTS

STABILITY

The readings given are the results at the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2006 to undertake the calibration presented in this certificate.

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page.

PROCEDURE

UIS procedure CP7.0.2

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

TEST EQUIPMENT USED

Asset No.	Description	Certificate No.	Expiry date
ID3156	Hart 1590 Super-Thermometer II	296385	21/MAR/2017
ID3051	Tinsley Std Resistor 25 ohms	UKAS 0591304	09/JUL/2018
ID3052	PRT (25 ohms)	UKAS 47039	14/JUL/2017
ID3289	PRT (25 ohms)	UKAS 47042	14/JUL/2017
ID3240	PRT (25 ohms)	UKAS 47013	12/JUL/2017
ID3245	PRT (25 ohms)	UKAS 47020	12/JUL/2017
ID3276	Fuko 1586-2588 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: **Universal Instrument Services Ltd.**

Date of Calibration: 3rd January 2017
 Date of Issue: 5th January 2017

Certificate Number: 00043450



0125



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 68 The Whittle Estate, Cambridge Road,
 Whelstone, Leicester LE8 6PA
 Tel: 0116 275 0123 Fax: 0116 275 0262
 Website: www.uiscal.com
 Email: sales@uiscal.co.uk

Page 1 of 2 Pages
 Approved Signatory

J. Bruce
 J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 561430

Manufacturer: Not Known
Description: T^m Thermocouple
Model: T-TYPE
Serial No.: Q39785
Asset No.: Q39785
Order No.: 263528
Date Received: 19th December 2016

CONDITION OF UNIT UNDER TEST

The Thermocouple was visually inspected prior to calibration YES/NO
Y

ADDITIONAL COMMENTS

STABILITY

The readings given are the results at the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate.

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page.

PROCEDURE

UIS procedure CP7.5.3

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

TEST EQUIPMENT USED

Asset No.	Description	Certificate No.	Expiry date
ID3156	Fair 1590 Super-Thermometer II	296365	21/MAR/2017
ID3051	Tinsley Std Resistor 25 ohms	UKAS 0501304	09/JUL/2016
ID3032	PRT (25 ohms)	UKAS 47033	14/JUL/2017
ID3289	PRT (25 ohms)	UKAS 47042	14/JUL/2017
ID3240	PRT (25 ohms)	UKAS 47014	12/JUL/2017
ID3246	PRT (25 ohms)	UKAS 47020	12/JUL/2017
ID3276	Fuke 1586-2508 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: **Universal Instrument Services Ltd.**

Date of Calibration: 3rd January 2017
 Date of Issue: 5th January 2017
 Certificate Number: 00049153



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 69 The Whittle Estate, Cambridge Road,
 Whatstones, Leicester LE6 6PA
 Tel: 0116 275 0123 Fax: 0116 275 0262
 Website: www.uiscal.com
 Email: sales@uiscal.co.uk

Page 1 of 2 Pages
 Approved Signatory

 J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 361433

Manufacturer: Not Known
Description: 'T' Thermocouple
Model: T-TYPE
Serial No: Q39786
Asset No: Q39786
Order No: 263526
Date Received: 15th December 2016

CONDITION OF UNIT UNDER TEST

The Thermocouple was visually inspected prior to calibration

YES/NO
 Y

ADDITIONAL COMMENTS

STABILITY

The readings given are the results at the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate.

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page.

PROCEDURE

UIS procedure CP7.5.3

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

TEST EQUIPMENT USED

<u>Asset No.</u>	<u>Description</u>	<u>Certificate No.</u>	<u>Expiry date</u>
ID3156	Hart 1580 Super- thermometer II	295385	21/MAR/2017
ID3051	Tinsley Std Resistor 25 ohms	UKAS 0391304	09/JUL/2016
ID3032	PRT (25 ohms)	UKAS 47039	14/JUL/2017
ID3269	PRT (25 ohms)	UKAS 47042	14/JUL/2017
ID3240	PRT (25 ohms)	UKAS 47018	12/JUL/2017
ID3245	PRT (25 ohms)	UKAS 47020	12/JUL/2017
ID3278	Fluke 1586-2588 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: **Universal Instrument Services Ltd.**

Date of Calibration: 3rd January 2017
 Date of Issue: 5th January 2017
 Certificate Number: UUI48452



0125



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 65 The Whittle Estate, Cambridge Road,
 Whetstone, Leicester LE8 5PA
 Tel: 0116 275 0123 Fax: 0116 275 0262
 Website: www.uiscal.com
 Email: sales@uiscal.co.uk

Page 1 of 2 Pages
 Approved Signatory

J. Bruce
 J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 351432

Manufacturer: Not Known
Description: T^h Thermocouple
Model: T-TYPE
Serial No: Q39767
Asset No: Q39787
Order No: 263528
Date Received: 19th December 2016

CONDITION OF UNIT UNDER TEST

The Thermocouple was visually inspected prior to calibration **YES/NO**
 Y

ADDITIONAL COMMENTS

STABILITY

The readings given are the results at the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate.

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page.

PROCEDURE

UIS procedure CP7.5.3

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

TEST EQUIPMENT USED

Asset No.	Description	Certificate No.	Expiry date
ID3166	Har. 1590 Super-Thermometer II	203385	21/MAR/2017
ID3061	Tinsley Std Resistor 25 ohms	UKAS 0391304	05/JUL/2016
ID3032	FRT (25 ohms)	UKAS 47039	14/JUL/2017
ID3289	FRT (25 ohms)	UKAS 47042	14/JUL/2017
ID3240	FRT (25 ohms)	UKAS 47018	12/JUL/2017
ID3276	PRT (25 ohms)	UKAS 47020	12/JUL/2017
ID3276	Fluke 1586-2688 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: **Universal Instrument Services Ltd.**

Date of Calibration: 3rd January 2017
 Date of Issue: 6th January 2017

Certificate Number: 00048451



0125



UNIVERSAL INSTRUMENT SERVICES Ltd.
 Unit 69 The Whittle Estate, Cambridge Road,
 Whetstons, Leicester LE16 6PA
 Tel: 0116 275 0123 Fax: 0116 275 0262
 Website: www.uiscal.com
 Email: sales@uiscal.co.uk

Page 1 of 2 Pages
 Approved Signatory

J. Bruce
 J. Bruce

Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE
Operator: JKB
Our Ref: 361431

Manufacturer: Not Known
Description: "T" Thermocouple
Model: T-TYPE
Serial No.: Q39788
Asset No.: Q30788
Order No.: 263528
Date Received: 19th December 2016

CONDITION OF UNIT UNDER TEST YES/NO
 The Thermocouple was visually inspected prior to calibration Y

ADDITIONAL COMMENTS

STABILITY

The readings given are the results at the time of calibration and do not carry any implication regarding the long term stability of the unit under test.

ACCREDITATIONS

UIS is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate

ENVIRONMENT

The instrument was calibrated in our laboratory with the ambient conditions stated on the results page

PROCEDURE

UIS procedure CP7.5.3

UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.


TEST EQUIPMENT USED

Asset No.	Description	Certificate No.	Expiry date
1D3155	Hart 1590 Super-Thermometer II	296365	21/MAR/2017
1D3051	Tinsley Std Resistor 25 ohms	UKAS 0381334	09/JUL/2018
1D3032	PRT (25 ohms)	UKAS 47039	14/JUL/2017
1D3239	PRT (25 ohms)	UKAS 47042	14/JUL/2017
1D3240	PRT (25 ohms)	UKAS 47018	12/JUL/2017
1D3245	PRT (25 ohms)	UKAS 47020	12/JUL/2017
1D3276	Fluke 1560-2588 Multiplexer	UKAS 47348	25/FEB/2017

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

Climatic Humidity Probe

31995 120

<p>CERTIFICATE OF CALIBRATION</p> <p>Issued by</p> <p>ABSOLUTE CALIBRATION LIMITED</p> <p>DATE OF ISSUE 18 August 2017 CERTIFICATE NUMBER 0428395</p>	
--	---

Page 1 of 2 Pages



Absolute Calibration Limited
 14 Murrills Estate, Portchester
 Hampshire, England. PO16 9RD
 Telephone 023-92321712
 Facsimile 023-92210034
 Service Facsimile 023-92327100
 www.absolute-cal.co.uk

Approved Signatory

 M Funnell
 S Whittingham
 D Kingswell
 G Mills
 A Francis

Description:	TEMPERATURE/HUMIDITY SENSOR AND ACTIVE ADAPTOR
Manufacturer:	ROTRONIC
Type Number:	HYGROCLIP & MOK-20-XX-010V-2
Serial Number:	60250059
Customer:	HORIBA MIRA LIMITED WATLING STREET NUNEATON WARWICKSHIRE

Instrument Receipt Date:	02 August 2017
Order Number:	271883
Customer Reference:	Q31995
Laboratory Temperature	(20.0 ± 3.0) °C
Laboratory Humidity	(55 ± 20) %rh
Calibration Procedure:	CP 112
Calibration Engineer	S PalaLendi
Calibration Date	18 August 2017


This Report Contains	Recorded results with no adjustments	<input checked="" type="checkbox"/>
	Pre and post adjustment results	<input type="checkbox"/>
	Post repair results	<input type="checkbox"/>
	Results recorded at Customer site	<input type="checkbox"/>

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognized national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

HM 8306

11 for Mira 6413/27055/201706

GMITB LWC

CERTIFICATE OF CALIBRATION		
Issued by ABSOLUTE CALIBRATION LIMITED		
DATE OF ISSUE	18 August 2017	CERTIFICATE NUMBER 0426394

Page 1 of 2 Pages



Absolute Calibration Limited
 14 Murrills Estate, Portchester
 Hampshire, England. PO16 9RD
 Telephone 023-92321712
 Facsimile 023-92210034
 Service Facsimile 023-92327100
 www.absolute-cal.co.uk

Approved Signatory

 M Funnell
 S Whittingham
 D Kingswell
 G Mills
 A Francis

Description:	TEMPERATURE/HUMIDITY SENSOR AND ACTIVE ADAPTOR
Manufacturer:	ROTRONIC
Type Number:	HYGROCLIP & MOK-20-XX-010V-2
Serial Number:	60250328
Customer:	HORIBA MIRA LIMITED WATLING STREET NUNEATON WARWICKSHIRE

Instrument Receipt Date:	02 August 2017
Order Number:	271883
Customer Reference:	Q31996
Laboratory Temperature	(20.0 ± 3.0) °C
Laboratory Humidity	(55 ± 20) %rh
Calibration Procedure:	CP 112
Calibration Engineer	S Patabendi
Calibration Date	18 August 2017

This Report Contains	Recorded results with no adjustments	<input checked="" type="checkbox"/>
	Pre and post adjustment results	<input type="checkbox"/>
	Post repair results	<input type="checkbox"/>
	Results recorded at Customer site	<input type="checkbox"/>

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and for 10 units of measurement realized at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

FM 5505

431497 CWO

CERTIFICATE OF CALIBRATION	
Issued by	
ABSOLUTE CALIBRATION LIMITED	
DATE OF ISSUE	19 January 2017
CERTIFICATE NUMBER	0416313



Page 1 of 2 Pages



Absolute Calibration Limited
 14 Murril's Estate, Portchester
 Hampshire, England, PO16 9RD
 Telephone 023-92321712
 Facsimile 023-92210034
 Service Facsimile 023-92327100
 www.absolute-cal.co.uk

Approved Signatory
n. Suresh
 M Tunnel
 S Whittingham
 D Kingswell
 G Mills
 A Francis

Description: THERMOHYGROMETER PROBE AND ACTIVE ADAPTOR
 Manufacturer: ROTRINIC
 Type Number: HYGROCLIP & MOK-20-XX-010V-2
 Serial Number: 5549S255
 Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE

Instrument Receipt Date: 17 January 2017
 Order Number: 284315
 Customer Reference: Q31997
 Laboratory Temperature: (20.0 ± 3.0) °C
 Laboratory Humidity: (55 ± 20) %rh
 Calibration Procedure: CP 112
 Calibration Engineer: S Patabendi
 Calibration Date: 19 January 2017

This Report Contains


Recorded results with no adjustments	<input checked="" type="checkbox"/>
Pre and post adjustment results	<input type="checkbox"/>
Post repair results	<input type="checkbox"/>
Results recorded at Customer site	<input type="checkbox"/>

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and for its units of measurement, realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

PM503

L:\working\121576503\040116

Q31998 CWØ

CERTIFICATE OF CALIBRATION		
Issued by		
ABSOLUTE CALIBRATION LIMITED		
DATE OF ISSUE	19 January 2017	CERTIFICATE NUMBER 0416612

Page 1 of 2 Pages



Absolute Calibration Limited

14 Murills Estate, Portchester
 Hampshire, England, PO16 9RD
 Telephone 023-92321712
 Facsimile 023-92210034
 Service Facsimile 023 92327100
 www.absolute-cal.co.uk

Approved Signatory

S. Patabendi
 M Funnell
 S Whittingham
 D Kingwell
 G Mills
 A Francis

Description: THERMOHYGROMETER PROBE AND ACTIVE ADAPTOR
 Manufacturer: ROTRINIC
 Type Number: HYGROCLIP & MOK-20-XX-010V-2
 Serial Number: 60250104
 Customer: HORIBA MIRA LIMITED
 WATLING STREET
 NUNEATON
 WARWICKSHIRE

Instrument Receipt Date: 17 January 2017
 Order Number: 284315
 Customer Reference: Q31998
 Laboratory Temperature: (20.0 ± 3.0)°C
 Laboratory Humidity: (55 ± 20) %rh
 Calibration Procedure: CP 112
 Calibration Engineer: S Patabendi
 Calibration Date: 18 January 2017

This Report Contains

Recorded results with no adjustments	<input checked="" type="checkbox"/>
Pre and post adjustment results	<input type="checkbox"/>
Post repair results	<input type="checkbox"/>
Results recorded at Customer site	<input type="checkbox"/>

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and for its units of mass, length realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing authority.

FM 06/05

U:\work\mg\121576503\0416612

32ch Thermocouple Amplifier

CALIBRATION CERTIFICATE

Issued by HORIBA MIRA Ltd

Issue date: 04 Jan 2017 **Cert No :** 30279191216



Page 1 of 5 Pages

Approved Signatory

P. MacLeod

Walling Street, Nuneaton
 Warwickshire, CV10 0TU, UK
 Telephone : +44 (0)24 7635 5225
 Facs mlie : +44 (0)24 7635 8225
<http://www.horiba-mira.com>

- Mark Pickering – Head of Instrument Calibration & Repair
- Philip Macleod – Supervisor – Instrument Calibration
- Dominic Mhandu - Metrologist

Client:	Test Operations FG	Manufacturer:	National Instru
Section:	Vehicle Env & Aero, Climatic Chamber	Model:	SCXI-1102
Address:	Walling Street	Description:	32ch Thermocouple Amplifier
	Nuneaton	Serial No:	122578B
	Warwickshire	Calibration Date:	04 Jan 2017
	--	Calibration Procedure:	QA2656/C/02
Client ID:	--	Equipment used:	Page 5
MIRA ID:	Q30279	Measurement Results:	Page 2-5
Date received:	18 Dec 2016	Measurement Uncertainty:	Page 2-5
Dallas ID:			

Condition of instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

The instrument was allowed to acclimatise in an environment of 20°C ± 2°C and 50%RH ± 25%RH, for a minimum of 12 hours before commencing the calibration. The electrical supply within the laboratory is 240 Volts ± 15 Volts and 50 Hz ± 0.5Hz with a total harmonic distortion of less than 3%.

This Instrument was calibrated by comparison with electrical measurement reference standards using a MIRA procedure which incorporates limits based on client requirements. These requirements are specified in document MC-15042003A



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Walling Street, Nuneaton, Warwickshire, CV10 0TU. Registered in England No. 0626252. VAT Registration number: GB10148494.

TC0032b Issue 20

Climatic Chamber Controller

CALIBRATION CERTIFICATE

Issued by HORIBA MIRA Ltd

Issue date: 05 Jan 2017 Cert No : 34000050117



Page 1 of 3 Pages

Approved Signatory

P. Macleod

Walling Street, Nuneaton
Warwickshire, CV10 0TU, UK.
Telephone : +44 (0)24 7635 6225
Facsimile : +44 (0)24 7635 8225
<http://www.horiba-mira.com>

- Mark Pickering – Head of Instrument Calibration & Repair
- Philip Macleod – Supervisor – Instrument Calibration
- Dominic Mhandu – Metallologist

Client:	Test Operations FG	Manufacturer:	MIRA
Section:	Vehicle Env & Aero, Climatic Chamber	Model:	M1003878
Address:	Walling Street Nuneaton Warwickshire --	Description:	Climatic Chamber Controller
Client ID:	--	Serial No:	I01523
MIRA ID:	Q34C0C	Calibration Date:	05 Jan 2017
Date received:	05 Jan 2017	Calibration Procedure:	QA5032/C/01
Dallas ID:		Equipment used:	Page 3
		Measurement Results:	Pages 2 & 3
		Measurement Uncertainty:	Page 5

Condition of Instrument: Used, in good condition

Within specification on receipt, at the points measured subject to the measurement uncertainty	Yes
Adjusted during calibration	No
Repaired prior to or during calibration	No
Within specification on completion, at the points measured subject to the measurement uncertainty	Yes

The reported values are the result of measurements taken at the time of calibration within the environment stated and do not carry any implication regarding the long term stability or environmental performance of the instrument. All measurements detailed within this Calibration Certificate relate only to the instrument detailed above on the dates specified.

This calibration was performed at the client site.

This instrument was calibrated by comparison with Temperature measurement reference standards using a MIRA procedure which incorporates limits based on client requirements as specified in document CW003032C15A.



The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of k=2, providing a level of confidence of approximately 95%.

Copyright 2017 HORIBA MIRA Ltd. All rights reserved subject to client contract. Information contained within this report may not be reproduced in any form without the prior written agreement of the CEO of HORIBA MIRA Ltd. All Measurements are Traceable to National Measurement Standards unless stated otherwise. MIRA Ltd, Registered office: Walling Street, Nuneaton, Warwickshire, CV10 0TU. Registered in England No: 8626352. VAT Registration number: GB 00146484.

TC0032b Issue 20